



# CERTIFICATE OF ANALYSIS

## OPA Diluent for Post-Column Carbamate Analysis

CB910

**Product:** OPA Diluent for Post Column Carbamate Analysis

**Part Number:** CB910

**Lot Number:** 101606

**Test Start Date:** 10 JAN 2011

**Expiration Date:** 07 JAN 2013


**Primary Application:** For preparation of post-column fluorescence reagent for carbamate pesticide analysis.

**Acceptance Criteria:**

- 1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C 9.186
- 2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
- 3. Twelve appropriate peaks in chromatogram (Y/N) Y

**Calibration Standard:** Carbamate Test Standard, PN 1700-0063, LN 908045

**Storage Conditions:** Dry at 8 to 25 °C

Analysis Certified by:  Date: 11 FEB 2011  
VP OPER Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE