



# CERTIFICATE OF ANALYSIS

## OPA Diluent for Post-Column Carbamate Pesticides Analysis

**CB910**

**Product:** OPA Diluent for Post Column analysis of Carbamate pesticides

**Part Number:** CB910

**Lot Number:** 101037

**Production Date:** 28 JAN 2011

**Expiration Date:** 27 JAN 2013

**Primary Application:** For preparation of post-column fluorescence reagent for Carbamate analysis.

### Acceptance Criteria:

- 1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C 9.144
- 2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
- 3. Twelve appropriate peaks in chromatogram (Y/N) Y

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063, **LN:** 908045

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:** *Jim Murphy* **Date:** 15 FEB 2011

*VP OPR*

Title

Rev. 8, 9/19/08

ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE