



## CERTIFICATE OF ANALYSIS

### OPA Diluent for Post-Column Carbamate Pesticides Analysis

CB910

**Product:** OPA Diluent for Post Column analysis of Carbamate pesticides

**Part Number:** CB910

**Lot Number:** 102021

**Expiration Date:** 14 Feb 2013

**Test Start Date:** 15 Feb 2011

**Primary Application:** For preparation of post-column fluorescence reagent for Carbamate analysis.

**Acceptance Criteria:**

1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C

9.135

2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)

Y

3. Twelve appropriate peaks in chromatogram (Y/N)

Y

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063, **LN:** 908045

**Storage Conditions:** Dry at 8°C to 25 °C

Analysis Certified by: \_\_\_\_\_

*Jim Murphy*  
VP Opr  
\_\_\_\_\_  
Title

Date: \_\_\_\_\_

17 Feb 2011

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE