



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-Column Carbamate Pesticides Analysis

CB910

Product: OPA Diluent for Post Column analysis of Carbamate pesticides

Part Number: CB910

Lot Number: 102029

Expiration Date: 22 Feb 2013

Test Start Date: 23 Feb 2011

Primary Application: For preparation of post-column fluorescence reagent for Carbamate analysis.

Acceptance Criteria:

- | | |
|---|--------------|
| 1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C | <u>9.121</u> |
| 2. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>Y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>Y</u> |

Sample: Carbamate Test Mixture, **PN:** 1700-0063, **LN:** 908045

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by:  **Date:** 02 MAR 2011

VP OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE