



## CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Carbamate Pesticides Analysis

CB910

Product: OPA Diluent for Post-column analysis of Carbamate pesticides

Part Number: CB910

Lot Number: 103023

Production Date: 15 MAR 2011

Expiration Date: 14 MAR 2013


Primary Application: Post-column analysis of Carbamate pesticides.

### Acceptance Criteria

1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N) 4
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) 4
3. Twelve appropriate peaks in chromatogram (Y/N) 4

Sample: Carbamate Test Mixture, PN: 1700-0063

Storage Conditions: Dry at 8 °C to 25 °C

Analysis Certified by:  Date: 21 MAR 2011  
VP Ops  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE