



# CERTIFICATE OF ANALYSIS

## OPA Diluent for Post-Column Carbamate Pesticides Analysis

**CB910**

**Product:** OPA Diluent for Post Column analysis of Carbamate pesticides

**Part Number:** CB910

**Lot Number:** 104007

**Production Date:** 06 Apr 2011

**Expiration Date:** 05 Apr 2013

**Primary Application:** Post-column analysis of Carbamate pesticides.

### Acceptance Criteria

- 1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N) 4
- 2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) 4
- 3. Twelve appropriate peaks in chromatogram (Y/N) 4

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at 8 °C to 25 °C

**Analysis Certified by:** *Jim Murphy* **Date:** 12 Apr 2011  
VP Ops  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE