



## CERTIFICATE OF ANALYSIS

### OPA Diluent for Post-column Carbamate Pesticides Analysis

**CB910**

**Product:** OPA Diluent for Post-column analysis of Carbamate pesticides

**Part Number:** CB910

**Lot Number:** 111037

**Production Date:** 29 Nov 2011

**Expiration Date:** 23 Nov 2013

**Primary Application:** Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Y  
Y  
Y

**Sample:** Carbamate Test Mixture, PN: 1700-0063

**Storage Conditions:** Dry at 8 °C to 25 °C

**Analysis Certified by:**

*Jim Lynch*  
VP OPR  
Title

**Date:**

01 Dec 2011

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE