



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Glyphosate and AMPA Analysis

GA104

Product: OPA Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA104

Lot Number: 104 013

Production Date: 07 APR 2011

Expiration Date: 06 APR 2013

Primary Application: Post-column analysis of glyphosate herbicide.

Acceptance Criteria:

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) Y

2. Glyphosate and AMPA peaks are visible and in order(Y/N) Y

3. Baseline noise (0 ± 25 mvolts) (Y/N) Y

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8 °C to 25 °C

Analysis Certified by: 

Date: 12 Apr 2011

Title

Rev.9, 04/05/2011

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE