



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

GA116

Product: Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA116

Lot Number: 105010

Production Date: 11 MAY 2011

Expiration Date: 11 MAY 2013

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

- 1. pH = 11.65 ± 0.1 @ 25 ± 0.4 °C (Y/N) y
- 2. Glyphosate & AMPA peaks are visible and in order (Y/N) y
- 4. Baseline noise (0 ± 25 mvolts) (Y/N) y

Sample: Glyphosate Test Mixture, **PN** 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: Jim Murphy **Date:** 18 MAY 2011

VP OPR

Title

ACCEPTED
PICKERING LABORATORIES, INC
QUALITY ASSURANCE