



## CERTIFICATE OF ANALYSIS

**OPA Diluent for Post-column Glyphosate and AMPA Analysis**

**GA104**

**Product:** OPA Diluent for Post-column Glyphosate and AMPA analysis

**Part Number:** GA104

**Lot Number:** 109015

**Production Date:** 21 Sep 2011

**Expiration Date:** 21 Sep 2013

**Primary Application:** Post-column analysis of Glyphosate herbicide.

**Acceptance Criteria:**

1. pH =  $10.5 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N) Y
2. Glyphosate and AMPA peaks are visible and in order(Y/N) Y
3. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y

**Sample:** Glyphosate Test Mixture, PN 1700-0080

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** *Jim Murphy* **Date:** 27 Sep 2011

*VPOPC*  
Title

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev.A, 04/18/2011

P:\Documentation\Chemical Documentation\GA104