



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Glyphosate and AMPA Analysis

GA104

Product: OPA Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA104

Lot Number: 201007

Production Date: 06 JAN 2012

Expiration Date: 05 JAN 2014

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:


1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Glyphosate and AMPA peaks are visible and in order(Y/N)
3. Baseline noise (0 ± 25 mvolts) (Y/N)

4
4
4

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:


V.P. OBER
Title

Date:

28 Feb 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE