



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Glyphosate and AMPA Analysis

GA104

Product: OPA Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA104

Lot Number: 512013

Production Date: 10 Dec 2015

Expiration Date: 10 Dec 2017

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Glyphosate and AMPA peaks are visible and in order(Y/N) | <u>y</u> |
| 3. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>y</u> |

Sample: Glyphosate Test Mixture, **PN** 1700-0080

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** 21 December 2015
President
Title