



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Glyphosate Analysis

GA116

Product: Hypochlorite Diluent for Glyphosate Analysis

Part Number: GA116

Lot Number: 007029

Test Start Date: 23 Jul 2010

Expiration Date: 23 Jul 2012

Primary Application: For preparation of post-column oxidizing reagent for Glyphosate herbicide analysis.

Acceptance Criteria:


- 1. pH = 11.65 ± 0.1 @ 25 ± 0.4 °C
 - 2. Glyphosate & AMPA peaks are visible and in order (Y/N)
 - 3. AMPA and Glyphosate area meet specifications (Y/N)
- NOTE:** Area ratios are affected by the hypochlorite added to GA116
- 4. Baseline noise (0 ± 25 mvolts) (Y/N)

Actual Values

11.690
Y
Y
Y

Calibration Standard: Glyphosate Test Standard, PN 1700-0080, LN 9/2031

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 29 Jul 2010
VP OPR
Title

Rev. 9, 09/25/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE