



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

GA116

Product: Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis
Part Number: GA116

Lot Number: 104015
Production Date: 11 APR 2011
Expiration Date: 11 APR 2013

Primary Application: For preparation of post-column oxidizing reagent for Glyphosate herbicide analysis.

Acceptance Criteria:

- 1. pH = 11.65 ± 0.1 @ 25 ± 0.4 °C (Y/N)
- 2. Glyphosate & AMPA peaks are visible and in order (Y/N)
- 4. Baseline noise (0 ± 25 mvolts) (Y/N)

 4

 4

 4

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: *[Signature]* **Date:** 18 Apr 2011
VPOPR

 Title

Rev. 10, 04/06/2011

ACCEPTED
 PICKERING LABORATORIES, INC.
 QUALITY ASSURANCE