



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

GA116

Product: Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA116

Lot Number: 201011

Production Date: 09 Jan 2012

Expiration Date: 06 Jan 2014

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:


1. pH = 11.65 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Glyphosate & AMPA peaks are visible and in order (Y/N)
3. Baseline noise (0 ± 25 mvolts) (Y/N)

Y
Y
Y

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by:


VP OPR
Title

Date:

28 Feb 2012

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE