



CERTIFICATE OF ANALYSIS

Potassium Eluant For Biogenic Amines Analysis

K600

Product: Potassium Eluant for Biogenic Amines Analysis

Part Number: K600

Lot Number: 207007

Production Date: 10 Jul 2012

Expiration Date: 10 Jul 2014

Primary Application: Eluant for ALKION Cation-exchange chromatography of polyamines.

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 6.00 ± 0.10 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Peak resolution meets specification (Y/N) | <u>y</u> |
| 3. Baseline: no peaks (Y/N) | <u>y</u> |

Storage Conditions: Dry at 8 to 25°C

Analysis Certified by: Jim Murphy **Date:** 19 July 2012
VP Operations
Title

Rev. 6, 21/21/2011