



# CERTIFICATE OF ANALYSIS

## Lithium Diluent for Amino Acid Analysis

Li220

**Product:** Lithium Diluent for Amino Acid Analysis

**Part Number:** Li220

**Lot Number:** 001010

**Test Start Date:** 15 JAN 2010

**Expiration Date:** 15 JAN 2011

**Primary Application:** Lithium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $2.36 \pm 0.01$  @  $25 \pm 0.4$  °C
2. Baseline: no peaks
3. Conductivity after adding lithium citrate  $6.08 \pm 0.01$  mS

2.356  
4  
4

**Storage Conditions:** Dry at 8 to 25 °C

Analysis Certified by: \_\_\_\_\_

Date: 09 Feb 2010

VP OPERATIONS  
/title

Rev. 5, 8/11/08

**ACCEPTED**

PICKERING LABORATORIES, INC  
QUALITY ASSURANCE