



## CERTIFICATE OF ANALYSIS

### Lithium Diluent for Amino Acid Analysis

**Li220**

**Product:** Lithium Diluent for Amino Acid Analysis

**Part Number:** Li220

**Lot Number:** 403017

**Production Date:** 17 Mar 2014

**Expiration Date:** 17 Mar 2016

**Primary Application:** Lithium-based amino acid analysis.

**Acceptance Criteria:**

- 1. pH = 2.36 ± 0.01 @ 25 ± 0.4 °C (Y/N) y
- 2. Baseline: no peaks (Y/N) y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** 02 Apr 2014  
President  
Title