



## CERTIFICATE OF ANALYSIS

### Lithium Diluent for Amino Acid Analysis

**Li220**

**Product:** Lithium Diluent for Amino Acid Analysis

**Part Number:** Li220

**Lot Number:** 002040

**Production Date:** February 24, 2020

**Expiration Date:** February 24, 2022

**Primary Application:** Lithium-based amino acid analysis.

**Acceptance Criteria:**

1. pH = 2.36 ± 0.01 @ 25 ± 0.4 °C (Y/N)

Y

2. Baseline: no peaks (Y/N)

Y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Rebecca Smith  
Vice President

Title

**Date:** 3/18/2020