



CERTIFICATE OF ANALYSIS

Lithium Diluent for Amino Acid Analysis

Li220

Product: Lithium Diluent for Amino Acid Analysis

Part Number: Li220

Lot Number: 202027

Production Date: 22 Feb 2012

Expiration Date: 21 Feb 2014


Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.36 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Baseline: no peaks (Y/N)

4
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: 
Title

Date: 05 MAR 2012

ACCEPTED
PICKERING LABORATORIES, INC
QUALITY ASSURANCE

Rev. 7, 04/08/2011

P:\Documentation\Chemical Documentation\Li220