



CERTIFICATE OF ANALYSIS

Lithium Eluant (pH 2.75) for Amino Acid Analysis

Li275

Product: Lithium Eluant pH 2.75 for Amino Acid Analysis

Part Number: Li275

Lot Number: 710001

Production Date: 02 Oct 2017

Expiration Date: 02 Oct 2019

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C (Y/N)

y

2. Resolution meets specifications:

• Threonine/Serine (Y/N)

y

• Asparagine/Glutamic acid (Y/N)

y

3. Elution time of L-Aspartic acid meets specifications (Y/N)

y

4. Baseline: no peaks (Y/N)

y

Sample: Lithium Calibration Standard, **PN:** 011006P

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy

Date: 11 Oct 2017

President

Title