



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis

Li275

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li275

Lot Number: 912029

Test Start Date: 16 Dec 09

Expiration Date: 16 Dec 10

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C
2. Resolution:
 - Threonine/Serine > 72% (Y/N)
3. Elution time of L-Aspartic acid meets specifications (Y/N)
4. Conductivity: 6.08 ± 0.01 mS

Actual Values

2.748
Y
Y
6.08

Calibration Standard:

Lithium Calibration Standard, PN 011006p, LN 909011

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: 

Date: 18 Jan 2010

VP Ope.

Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 6, 8/11/08