



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis

Li275LC

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li275LC

Lot Number: 008034

Test Start Date: 25 AUG 2010

Expiration Date: 25 AUG 2011

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C
2. Resolution:
 - Threonine/Serine > 72% (Y/N)
3. Elution time of L-Aspartic acid meets specifications (Y/N)
4. Conductivity: 6.20 ± 0.01 mS

Actual Values

2.756

Y

Y

6.19

Calibration Standard:

Lithium Calibration Standard, PN 011006p, LN 909011

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Murphy
VP DPR
Title

Date: 30 Aug 2010

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 1, 05/03/10