



CERTIFICATE OF ANALYSIS

Lithium Eluant (pH 2.75) for Amino Acid Analysis

Li275LC

Product: Lithium Eluant pH2.75 for Amino Acid Analysis

Part Number: Li275LC

Lot Number: 110003

Production Date: 07 Oct 2011

Expiration Date: 05 Oct 2013


Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

- 1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C (Y/N) 4
- 2. Resolution meets specifications: 4
 - Threonine/Serine (Y/N)
 - Asparagine/ Glutamic acid (Y/N)
- 3. Elution time of L-Aspartic acid meets specifications (Y/N) 4
- 4. Baseline no peaks (Y/N) 4

Sample: Lithium Calibration Standard, PN: 011006P

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 19 Oct 2011
VP OPR
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE