



CERTIFICATE OF ANALYSIS

Lithium Eluant (pH 2.75) for Amino Acid Analysis

Li275LC

Product: Lithium Eluant pH2.75 for Amino Acid Analysis

Part Number: Li275LC

Lot Number:

111039

Production Date:

30 Nov 2011

Expiration Date:

28 Nov 2013

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:


1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Resolution meets specifications:
 - Threonine/Serine (Y/N)
 - Asparagine/Glutamic acid (Y/N)
3. Elution time of L-Aspartic acid meets specifications (Y/N)
4. Baseline: no peaks (Y/N)

Y
Y
Y
Y
Y

Sample: Lithium Calibration Standard, PN: 011006P

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:


V.P. Patel
Title

Date:

05 Dec 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE