



CERTIFICATE OF ANALYSIS

Lithium Diluent for Amino Acid Analysis

Li220

Product: Lithium Diluent for Amino Acid Analysis

Part Number: Li220

Lot Number: 008035

Test Start Date: 27 Aug 2010

Expiration Date: 27 Aug 2011

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.36 ± 0.01 @ 25 ± 0.4 °C
2. Baseline: no peaks
3. Conductivity after adding lithium citrate 6.08 ± 0.01 mS

2.362
4
6.08

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim P. [Signature]
VP Ops
Title

Date: 31 Aug 2010

Rev. 5, 8/11/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE