



CERTIFICATE OF ANALYSIS

Lithium Eluant (pH 2.75) for Amino Acids Analysis

Li280

Product: Lithium Eluant pH 2.75 for Amino Acids Analysis

Part Number: Li280

Lot Number: 201035

Production Date: 27 JAN 2012

Expiration Date: 26 JAN 2014

Primary Application: Lithium-based amino acids analysis.

Acceptance Criteria:

1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Resolution meets specifications:
 - Threonine/Serine (Y/N)
 - Asparagine/Glutamic acid (Y/N)
3. Elution time of L-Aspartic acid meets specifications (Y/N)
4. Baseline: no peaks (Y/N)

Y
Y
Y
Y
Y

Sample: Lithium Calibration Standard, PN: 011006P

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

[Signature]
Title

Date:

28 FEB 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE