



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis

Li292

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li292

Lot Number: 007011

Test Start Date: 08 Jul 2010

Expiration Date: 08 Jul 2011

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

Actual Values

1. pH = 2.91 ± 0.01 @ 25 ± 0.4 °C
2. Resolution:
 - Threonine/Serine $\geq 80\%$
 - Asparagine/ Glutamic Acid $\geq 50\%$
 - Glycine, Alanine, Citrulline $\geq 90\%$
3. Elution time:
 - L-aspartic acid = 7.9 ± 0.5 min.
4. Baseline: no peaks (Y/N)
5. Conductivity: 7.82 – 7.85 mS

2.903

4

4

4

4

4

7.84

Calibration Standard:

Lithium Calibration Standard, PN 1700-0170 LN 001047

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: [Signature] Date: 15 Jul 2010

VP OPR
Title

Rev. 9, 8/11/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE