



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis

Li365

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li365

Lot Number: 101021

Test Start Date: 21 JAN 2011

Expiration Date: 21 JAN 2013

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 3.65 ± 0.01 @ 25 ± 0.4 °C
2. Conductivity: 14.41 ± 0.01 mS @ 25 ± 0.4 °C
3. Resolution:
4. Elution Time
5. Baseline: No Peaks (Y/N)

Actual Values

<u>3.649</u>
<u>14.41</u>
<u>-</u>
<u>-</u>
<u>4</u>

Calibration Standard:

Lithium Calibration Standard, PN 1700-0170 LN 903003

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  Date: 15 Feb 2011

Title

Rev. 9, 7/23/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE