



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acids Analysis

Li365

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li365

Lot Number: 201021

Production Date: 20 Jan 2012

Expiration Date: 19 JAN 2014

Primary Application: Lithium-based amino acids analysis.

Acceptance Criteria:

- 1. pH = 3.65 ± 0.01 @ 25 ± 0.4 °C (Y/N) 4
- 2. Resolution meets specifications: 4
 - Glycine, Alanine, Citrulline, α-Amino-n-butyric acid and Valine (Y/N) 4
 - Cystathionine/Isoleucine (Y/N) 4
- 3. Elution time of β-Alanine meets specification (Y/N) 4
- 4. Baseline: no peaks (Y/N) 4

Calibration Standard: Lithium Calibration Standard, PN: 1700-0170

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: *Jim Murphy* **Date:** 29 Feb 2012
VP OPR
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 12, 10/05/2011

P:\Documentation\Chemical Documentation\Li365