



CERTIFICATE OF ANALYSIS

Lithium Eluant (pH3.65) for Amino Acids Analysis

Li365

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li365

Lot Number:

203035

Production Date:

30 MAR 2012

Expiration Date:

29 MAR 2014

Primary Application: Lithium-based amino acids analysis.

Acceptance Criteria:

1. pH = 3.65 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Resolution meets specifications:
 - Glycine, Alanine, Citrulline, α -Amino-n-butyric acid and Valine (Y/N)
 - Cystathionine/Isoleucine (Y/N)
3. Elution time of β -Alanine meets specification (Y/N)
4. Baseline: no peaks (Y/N)

Y
Y
Y
Y
Y

Calibration Standard: Lithium Calibration Standard, PN: 1700-0170

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

[Signature]
Title

Date:

06 Apr 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 12, 10/05/2011

P:\Documentation\Chemical Documentation\Li365