



## CERTIFICATE OF ANALYSIS

### Lithium Eluant for Amino Acid Analysis Li750

**Product:** Lithium Eluant for Amino Acid Analysis

**Part Number:** Li750

**Lot Number:** 304023

**Production Date:** 25 April 2013

**Expiration Date:** 25 April 2014

**Primary Application:** Lithium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $7.50 \pm 0.15$  @  $25 \pm 0.4$  °C (Y/N)

y

2. Resolution:

- All peaks in the Lithium Calibration standard are clearly resolved (Y/N)

y

3. Elution time of Histidine meets specifications (Y/N)

y

4. Baseline: no peaks (Y/N)

y

**Sample:** Lithium Calibration Standard, **PN:** 011006P

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy  
President

Title

**Date:** 28 May 2013