



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis
Part Number: Li750
Lot Number: 008008
Test Start Date: 04 Aug 2010
Expiration Date: 03 Aug 2011

Primary Application: Lithium amino acid analysis

Acceptance Criteria:

1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C
2. Conductivity: 7.73 ± 0.01 mS
3. Resolution:
 - All peaks in the Lithium Calibration Standard must be clearly resolved. (Y/N)
4. Elution time: (For Agilent 1100 HPLC System)
 - Histidine = 97.4 ± 0.5 min.
5. Baseline: no peaks (Y/N)

Actual Values:

7.603
7.73
Y
Y
Y

Calibration Standard:

Lithium Calibration Standard, PN 011006p, LN 909011

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

[Signature]
VP OPR

Title

Date: _____

06 Aug 2010