



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis
Part Number: Li750
Lot Number: 111017
Production Date: 02 Dec 2011
Expiration Date: 30 Nov 2012

Primary Application: Lithium amino acid analysis

Acceptance Criteria:

1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C (Y/N)
2. Resolution:
 - All peaks in the Lithium Calibration Standard must be clearly resolved. (Y/N)
3. Elution time of Histidine meets specification (Y/N)
4. Baseline: no peaks (Y/N)

Y
Y
Y
Y

Calibration Standard:

Lithium Calibration Standard, PN: 011006P,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

[Signature]
VP DPR
Title

Date: 05 Dec 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 10, 04/08/2011