



## CERTIFICATE OF ANALYSIS

### Lithium Eluant for Amino Acid Analysis Li750

**Product:** Lithium Eluant for Amino Acid Analysis

**Part Number:** Li750

**Lot Number:** 201064

**Production Date:** 05 JAN 2012

**Expiration Date:** 04 JAN 2013

**Primary Application:** Lithium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $7.50 \pm 0.15$  @  $25 \pm 0.4$  °C (Y/N)

2. Resolution:

- All peaks in the Lithium Calibration standard are clearly resolved (Y/N)

3. Elution time of Histidine meets specifications (Y/N)


4. Baseline: no peaks (Y/N)

Y  
Y  
Y  
Y

**Sample:** Lithium Calibration Standard, PN: 011006P

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**

  
Title

**Date:**

17 FEB 2012

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE