



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li750

Lot Number: 708032

Production Date: 15 Aug 2017

Expiration Date: 15 Aug 2018

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Resolution: | |
| • All peaks in the Lithium Calibration standard are clearly resolved (Y/N) | <u>y</u> |
| 3. Elution time of Histidine meets specifications (Y/N) | <u>y</u> |
| 4. Baseline: no peaks (Y/N) | <u>y</u> |

Sample: Lithium Calibration Standard, **PN:** 011006P

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** August 17, 2017
President
Title