



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.15) For Amino Acids Analysis

Na315

Product: Sodium Eluant pH 3.15 for Amino Acid Analysis

Part Number: Na315

Lot Number: 105007

Production Date: 03 MAY 2011

Expiration Date: 02 MAY 2013

Primary Application: Sodium eluant for Sodium-based amino acid analysis.

Acceptance Criteria:

1. Retention time of L-aspartic acid meets specifications (Y/N)
2. Resolution of Thr/Ser \geq 93% (Y/N)
3. Height of ammonia plateau/height of ammonia peak \leq 9% (Y/N)
4. No peaks in baseline (Y/N)

Y
Y
Y
Y

Sample: Collagen hydrolysate, PN 012506C,

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: _____

Date: 10 MAY 2011

Title: VP QPR

ACCEPTED
PICKERING LABORATORIES, INC
QUALITY ASSURANCE

Rev. 11, 2/24/2011