



## CERTIFICATE OF ANALYSIS

### Sodium Eluant (pH 3.15) For Amino Acids Analysis

**Na315**

**Product:** Sodium Eluant pH 3.15 for Amino Acid Analysis

**Part Number:** Na315

**Lot Number:** 409033

**Production Date:** 29 Sep 2014

**Expiration Date:** 29 Sep 2016

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N) y

2. Resolution meets specifications:  
• Threonine/Serine (Y/N) y

3. Elution time of L-Aspartic acid meets specifications (Y/N) y

4. Baseline: no peaks (Y/N) y

**Sample:** Collagen hydrolysate, **PN:** 012506C

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** 02 Oct 2014  
President  
Title