



**CERTIFICATE OF ANALYSIS**  
**Sodium Eluant for Amino Acid Analysis**  
**Na328**

**Product:** Sodium Eluant for Amino Acid Analysis  
**Part Number:** Na328  
**Lot Number:** 007027  
**Test Start Date:** 23 Jul 2010

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $3.28 \pm 0.01$  @  $25 \pm 0.4$  °C
2. Retention time of L-aspartic acid =  $8.5 \pm 0.5$  minutes.
3. No peaks on baseline.

**Actual Values**

3.284  
4  
4

**Calibration Standard:** Protein Hydrolysate, PN 012506H, LN 709037

**Analysis Certified by:** *[Signature]* **Date:** 27 Jun 2010  
VP OPR  
 Title

**ACCEPTED**

PICKERING LABORATORIES, INC.  
 QUALITY ASSURANCE

Rev. 5, 10/26/06