



## CERTIFICATE OF ANALYSIS

### Sodium Eluant (pH 3.28) For Amino Acids Analysis

**Na328**

**Product:** Sodium Eluant pH 3.28 for Amino Acid Analysis

**Part Number:** Na328

**Lot Number:** 205019

**Production Date:** 21 May 2012

**Expiration Date:** 21 May 2014

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH = $3.28 \pm 0.01$ @ $25 \pm 0.4$ °C (Y/N)               | <u>y</u> |
| 2. Elution time of L-Aspartic acid meets specifications (Y/N) | <u>y</u> |
| 3. Baseline: no peaks (Y/N)                                   | <u>y</u> |

**Sample:** Protein Hydrolysate, **PN:** 012506H

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** 29 May 2012  
VP Operations  
Title