



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.28) For Amino Acids Analysis

Na328

Product: Sodium Eluant pH 3.28 for Amino Acid Analysis

Part Number: Na328

Lot Number: 803001

Production Date: 01 Mar 2018

Expiration Date: 01 Mar 2020

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

- | | |
|---|----------|
| 1. pH = 3.28 ± 0.01 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Elution time of L-Aspartic acid meets specifications (Y/N) | <u>y</u> |
| 3. Resolution of Threonine-Serine meets specification (Y/) | <u>y</u> |
| 4. Baseline: no peaks (Y/N) | <u>y</u> |

Sample: Protein Hydrolysate, **PN:** 012506H

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** March 7, 2018
President
Title