



CERTIFICATE OF ANALYSIS
Sodium Eluant for Amino Acid Analysis

Na425

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na425

Lot Number: 201010

Production Date: 10 JAN 2012

Expiration Date: 09 JAN 2014

Acceptance Criteria:

1. pH = 4.25 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Resolution meets specifications:
 - Alanine/Cystine (Y/N)
 - Isoleucine/Leucine (Y/N)
3. Elution time of Valine and Methionine meets specification (Y/N)
4. Baseline: no peaks (Y/N)

Y
Y
Y
Y

Calibration Standard: Protein Hydrolysate Standard, PN: 012506H

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

[Signature]
V.P. OBR
Title

Date:

05 MAR 2012

ACCEPTED

Rev. 7, 12/20/2011

PICKERING LABORATORIES, INC
QUALITY ASSURANCE