



CERTIFICATE OF ANALYSIS
Sodium Eluant pH 4.25 for Amino Acid Analysis

Na425

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na425

Lot Number: 408031

Production Date: 26 Aug 2014

Expiration Date: 26 Aug 2016

Acceptance Criteria:

1. pH = 4.25 ± 0.01 @ 25 ± 0.4 °C (Y/N)

y

2. Resolution meets specifications:

• Alanine/Cystine (Y/N)

y

• Isoleucine/Leucine (Y/N)

y

3. Elution time of Valine and Methionine meets specification (Y/N)

y

4. Baseline: no peaks (Y/N)

y

Calibration Standard: Protein Hydrolysate Standard, **PN:** 012506H

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

Jim Murphy

Date:

04 Sep 2014

President

Title

Rev. A, 10/24/12

P:\Documentation\Chemical Documentation\Na425