



CERTIFICATE OF ANALYSIS
Sodium Eluant for Amino Acid Analysis
Na640

Product: Sodium Eluant for Amino Acid Analysis
Part Number: Na640
Lot Number: 201006
Production Date: 10 JAN 2012
Expiration Date: 09 JAN 2013

Acceptance Criteria:

- 1. pH = 6.40 ± 0.01 @ 25 ± 0.4 °C (Y/N) Y
- 2. Resolution meets specifications:
 - Lysine, Tryptophan, and Ammonia (Y/N) Y
- 3. Elution time of Histidine and Arginine meets specification (Y/N) Y
- 4. Baseline: no peaks (Y/N) Y

Calibration Standard: Protein Hydrolysate Standard, PN: 012506H

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: *Jim Murphy* **Date:** 05 MAR 2012
VP OPR
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev.8, 12/12/2011