



CERTIFICATE OF ANALYSIS
Sodium Eluant pH640 for Amino Acid Analysis
Na640

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na640

Lot Number: 2109036

Production Date: September 23, 2021

Expiration Date: September 23, 2022

Acceptance Criteria:

- | | |
|---|----------|
| 1. pH = 6.40 ± 0.01 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. Resolution meets specifications: | |
| • Lysine, Tryptophan, and Ammonia (Y/N) | <u>Y</u> |
| 3. Elution time of Histidine and Arginine meets specification (Y/N) | <u>Y</u> |
| 4. Baseline: no peaks (Y/N) | <u>Y</u> |

Calibration Standard: Protein Hydrolysate Standard, **PN:** 012506H

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** October 1, 2021
President
Title