



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Amino Acids Analysis

OD104

Product: OPA Diluent for Post-column Amino Acids analysis

Part Number: OD104

Lot Number: 105011

Production Date: 12 MAY 2011

Expiration Date: 12 MAY 2013

Primary Application: Post-column analysis of amino acids.

Acceptance Criteria:

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Peaks visible and eluted in order (Y/N)
4. Baseline noise (0 ± 25 mvolts) (Y/N)

Y
Y
Y

Sample: Glyphosate Test Mixture, PN 1700-0080,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

Jim Murphy
VP QCR

Date:

18 MAY 2011

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE