



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Lithium-based Amino Acids Analysis

**RG003**

**Product:** Lithium Column Regenerant

**Part Number:** RG003

**Lot Number:** 410005

**Production Date:** 03 Oct 2014

**Expiration Date:** 03 Oct 2016

**Primary Application:** For column regeneration in lithium-based amino acids analysis

**Acceptance Criteria:**

- 1. pH =  $12.60 \pm 0.20$  @  $25 \pm 0.4$  °C (Y/N) y
- 2. Conductivity meets specifications (Y/N) y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** 09 Oct 2014  
President  
Title