



CERTIFICATE OF ANALYSIS

Column Regenerant for Lithium-based Amino Acids Analysis

RG003

Product: Lithium Column Regenerant

Part Number: RG003

Lot Number: 611013

Production Date: 08 Nov 2016

Expiration Date: 08 Nov 2018

Primary Application: For column regeneration in lithium-based amino acids analysis

Acceptance Criteria:

1. pH = 12.60 ± 0.20 @ 25 ± 0.4 °C (Y/N)

y

2. Conductivity meets specifications (Y/N)

y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy

Date: November 16,
2016

President

Title