



CERTIFICATE OF ANALYSIS

Regenerant for Sodium System RG011 / 1700-1120

Product: Regenerant for Sodium System

Part Number: RG011 / 1700-1120

Lot Number: 012038

Test Date: 22 DEC 2010

Expiration Date: 22 DEC 2012

Primary Application: Sodium-based amino acid analysis

Acceptance Criteria:


1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C
2. Conductivity = 31.3 ± 0.7 @ 25 ± 0.4 °C

Actual Values:

12.67
31.5

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____



Date: 18 JAN 2011

VP QAR

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 9, 7/09/08