



CERTIFICATE OF ANALYSIS

Regenerant for Sodium System RG011 / 1700-1120

Product: Regenerant for Sodium System

Part Number: RG011 / 1700-1120

Lot Number: 106 029

Production Date: 23 JUN 2011

Expiration Date: 23 JUN 2013

Primary Application: Sodium-based amino acid analysis

Acceptance Criteria:

1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C (Y/N)
2. Conductivity = 31.3 ± 0.7 @ 25 ± 0.4 °C (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

Jim [Signature]
V.P. Ops

Title

Date:

28 JUN 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE